

Chloride

ALTERNATE NAMESCl⁻**SPECIMEN TYPE**

Serum or heparin plasma

STORAGE

Refrigerate

ROUTINE / STAT T.A.T.

24 hours / 30 mins

REFERENCE RANGE0-30 days: 98 -113 mmol/L
>30 days: 98-107 mmol/L**STABILITY**

1 week refrigerated

SPECIAL INSTRUCTIONS